


Microwave Digestion System

Make of microwave digestion system: CEM Corporation.

Model: Mars 6

Type of analysis possible:

- Sample digestion for qualitative Analysis
- Sample digestion for quantitative Analysis



- Selection of different instrument pre-program method.
- Selection of different user based method.
- Method can be development by changing the ramp temperature, ramp time and holding time.

Specification:

- Closed vessel microwave digestion at high temperature and pressure.
- Minimum samples size: 0.1 gm

Contact us:

Mobile No.

Email ID: Pankaj.Mishra@icar.gov.in

User Instructions :

- Solid sample should be grinded properly.
- Source of the sample should be mentioned.
- Samples should not be toxic or hazardous